

Approvals and conformities

ASME
RCC-M

Manufacturer : NDT-Europa (NL)**Description / Application :**

Translucent semi water-based blue gel.

It was specially developed in order to make ultrasonic testing on yards and wide surfaces easier.

OG-1 gel characteristic ensures a good adhesion on pieces to be tested, whatever their position.

OG-1 excellent wetting enables a very good acoustic transmission even on parts fat and not perfectly degreased.

OG-1 couplant has no corrosive effect on all metals.

OG-1 has no effect too on paints, plastic and rubber.

However it shall not be used on composites before repair.

The normal temperature range is 0°C to 50°C, the use is possible at higher temperature but the gel liquefies gradually.

The viscosity increases with evaporation of water.

Companion products: /

DIRECTIONS FOR USE

Apply on the part to be tested and then move US probe.

After testing, cleaning can be carried out with water or solvents. It is easy to wipe with cloth.

For a thorough cleaning, the residual presence of gel with blue fluorescent tracer incorporated can be controlled under UV-A.

The gel can be diluted with water if necessary, reducing corrosion properties accordingly.

TECHNICAL CHARACTERISTICS

Appearance translucent blue gel
Fluorescence blue

PRECAUTIONS FOR USE AND STORAGE

Transport / Handling: Refer to Material Safety Data Sheet (MSDS).

Storage : Temperature range: 5°C à 40°C

Keep packaging closed after taking out some of the product.

This technical data sheet replaces and cancels the previous one.

The above details have been compiled to the best of our knowledge. They have, however, an indicative value only and we therefore make no warranties and assume no liability in connection with any use of this information, particularly if a third party's rights are affected by the use of our products. The above information has been compiled based upon tests carried out by SOCOMORE. All data is subject to change as Socomore deems appropriate. The data given is not intended to substitute for any testing you must conduct in order to determine the suitability of the product for your particular purposes. Please check your local legislation applicable to the use of this product. Should you need any further information please contact us.